

IEC**IECEE**

®

TM

Ref. Certif. No.

DE 2-038017

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME**CB TEST CERTIFICATE**

Product	Genetic Sequencers
Name and address of the applicant	Illumina, Inc. 5200 Illumina Way San Diego CA 92122, USA
Name and address of the manufacturer	Illumina, Inc. 5200 Illumina Way San Diego CA 92122, USA
Name and address of the factory	Illumina, Inc. 25861 Industrial Blvd. Hayward CA 94545, USA
Ratings and principal characteristics	AC 200-240V, 50/60Hz; 2700W; Class I
Trademark (if any)	illumina
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	NovaSeq™ X, NovaSeq™ X Plus
Additional information (if necessary may also be reported on page 2)	
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010+A1 IEC 61010-2-010:2019 IEC 61010-2-081:2019 IEC 60825-1:2014 for national differences see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	US22N6FB 001

This CB Test Certificate is issued by the National Certification Body

**TÜVRheinland**®

TÜV Rheinland LGA Products GmbH
Tillystr. 2, 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web : www.tuv.com

Date: 2022-11-16

Signature:

Dipl.-Ing. M. Glagla